

Quantitative Analysis Results Summary Report

Batch Path	D:\MassHunter\GCMS\1\data\02sep22\QuantResults\02sep22.batch.bin	Analyst Name	DESKTOP-R0LAL7E\analyst
Analysis Time	9/6/2022 9:40 AM	Reporter Name	DESKTOP-R0LAL7E\analyst
Report Time	9/22/2022 10:57:19 AM	Batch State	Processed
Last Calib Update	9/6/2022 9:10 AM	Quant Report Version	10.1
Quant Batch Version	10.1		

Sequence Table

Data File	sample Name	Sample Type	Vial Position	Inj Vol	Level	Acq Method File
tq090202.D	2.0 uL #3318-24-0.5 CCV	CC	2	0	L6	tq22m0831
tq090204.D	2.0 uL #3318-23-0.5 LCS	QC	4	0	L6	tq22m0831
tq090206.D	2.0 uL #3257-128-0.5A MDL Blank	Sample	6	0		tq22m0831

Quantitation Results

Compound: 4:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	11.742	12219	33021	0.3700	0.8894	1.0000	88.9
tq090204.D	QC	6:2 FTOH-C13	11.742	10481	27837	0.3765	0.9050	1.0000	90.5
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 5:2sFTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	11.806	5368	33021	0.1626	0.9216	1.0000	92.2
tq090204.D	QC	6:2 FTOH-C13	11.800	4776	27837	0.1716	0.9726	1.0000	97.3
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 7:2sFTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	13.111	10568	33021	0.3200	0.8473	1.0000	84.7
tq090204.D	QC	6:2 FTOH-C13	13.111	9332	27837	0.3352	0.8875	1.0000	88.8
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 6:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	13.145	21051	33021	0.6375	0.8269	1.0000	82.7
tq090204.D	QC	6:2 FTOH-C13	13.137	19091	27837	0.6858	0.8895	1.0000	88.9
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 8:2 FTOH-C13

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	14.342	9471	33021	0.2868	0.9282	1.0000	92.8
tq090204.D	QC	6:2 FTOH-C13	14.342	9085	27837	0.3264	1.0562	1.0000	105.6
tq090206.D	Sample	6:2 FTOH-C13	14.348	8369	26464	0.3162	1.0234		

Compound: 8:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	14.348	6304	33021	0.1909	0.8493	1.0000	84.9

Quantitative Analysis Results Summary Report

Compound: 8:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090204.D	QC	6:2 FTOH-C13	14.342	5907	27837	0.2122	0.9440	1.0000	94.4
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 10:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	15.416	2238	33021	0.0678	0.9015	1.0000	90.1
tq090204.D	QC	6:2 FTOH-C13	15.410	2158	27837	0.0775	1.0311	1.0000	103.1
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: 12:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	16.397	8383	33021	0.2539	0.8910	1.0000	89.1
tq090204.D	QC	6:2 FTOH-C13	16.391	6614	27837	0.2376	0.8339	1.0000	83.4
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: NMeFOSA

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	18.601	6158	33021	0.1865	1.0392	1.0000	103.9
tq090204.D	QC	6:2 FTOH-C13	18.593	4031	27837	0.1448	0.8070	1.0000	80.7
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		

Compound: NEtFOSA

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq090202.D	CC	6:2 FTOH-C13	18.870	27839	33021	0.8431	1.0498	1.0000	105.0
tq090204.D	QC	6:2 FTOH-C13	18.862	18190	27837	0.6535	0.8137	1.0000	81.4
tq090206.D	Sample	6:2 FTOH-C13			26464		ND		