

Quantitative Analysis Results Summary Report

Batch Path	D:\MassHunter\GCMS\1\data\30sep22\QuantResults\30sep22.batch.bin	Analyst Name	DESKTOP-R0LAL7E\analyst
Analysis Time	10/3/2022 10:17 AM	Reporter Name	ES-AREA1\US32_USR_INS22923
Report Time	10/20/2022 2:37:17 PM	Batch State	Processed
Last Calib Update	10/3/2022 8:55 AM	Quant Report Version	10.1
Quant Batch Version	10.1		

Sequence Table

Data File	sample Name	Sample Type	Vial Position	Inj Vol	Level	Acq Method File
tq100303.D	2.0 uL #3318-33-0.5 LCS/ICV	QC	3	0	L6	tq22m0930markes

Quantitation Results

Compound: 4:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	11.819	35945	72387	0.4966	0.9034	1.0000	90.3

Compound: 5:2sFTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	11.882	14541	72387	0.2009	0.8489	1.0000	84.9

Compound: 7:2sFTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	13.183	28382	72387	0.3921	0.8239	1.0000	82.4

Compound: 6:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	13.209	62904	72387	0.8690	0.8606	1.0000	86.1

Compound: 8:2 FTOH-C13

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	14.406	24423	72387	0.3374	0.9669	1.0000	96.7

Compound: 8:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	14.407	18754	72387	0.2591	0.9326	1.0000	93.3

Compound: 10:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	15.477	6350	72387	0.0877	0.8395	1.0000	84.0

Compound: 12:2 FTOH

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	16.452	22213	72387	0.3069	0.8083	1.0000	80.8

Compound: NMeFOSA

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	18.654	18527	72387	0.2559	0.9125	1.0000	91.3

Quantitative Analysis Results Summary Report

Compound: NEtFOSA

Data File	Sample Type	ISTD	RT	Resp	ISTD Resp	Resp Ratio	Final Conc	Exp. Conc	Accuracy
tq100303.D	QC	6:2 FTOH-C13	18.924	77943	72387	1.0768	0.8771	1.0000	87.7